
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000051
9:34 06/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:31
STERILIZING 3:35
136.0 C 220kPa 3:35
136.0 C 221kPa 4:05
136.1 C 222kPa 4:35
136.5 C 221kPa 5:05
135.7 C 220kPa 5:35
136.3 C 223kPa 6:05
136.3 C 224kPa 6:35
136.1 C 224kPa 7:05
136.1 C 224kPa 7:05
VENTING 7:05
AIR DRYING 7:55
CYCLE COMPLETE 22:55

Digital Signature #
9B78CD8CEF506ACD
