
Midmark M11 - v1.0.5
Total Cycles: 962

Name: RITTER2
Sterilizer ID: V2313228
04/06/2022 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 187.0 F 0.3
4:00 198.3 F 2.9
6:00 213.4 F 8.1
8:00 228.9 F 13.5
10:00 233.9 F 10.3
12:00 234.5 F 8.4
14:00 240.0 F 10.4
16:00 248.7 F 15.0
18:00 261.1 F 22.1
20:00 272.3 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.3 F 29.7
2:00 272.3 F 29.2
3:00 271.8 F 29.3
4:00 272.5 F 29.5

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 212.7 F 0.0
10:00 224.6 F 0.0
15:00 229.4 F 0.0
20:00 226.1 F 0.0
25:00 225.5 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 20:21
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

04/06/2022 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____