

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001837  
11:39 12/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:52  
PRESSURIZING 1:55  
STERILIZING 2:59  
136.4 C 224kPa 2:59  
136.4 C 226kPa 3:29  
136.4 C 225kPa 3:59  
136.2 C 224kPa 4:29  
136.3 C 225kPa 4:59  
136.3 C 225kPa 5:29  
136.2 C 223kPa 5:59  
136.4 C 223kPa 6:29  
136.4 C 223kPa 6:29  
VENTING 6:30  
AIR DRYING 7:21  
CYCLE COMPLETE 11:35

-----  
Digital Signature #  
13FB63F27ADD9824  
-----