
Midmark M11 - v1.0.0
Total Cycles: 3863

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2023 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.7 F 0.0
2:00 165.0 F 2.3
4:00 179.1 F 5.8
6:00 197.4 F 10.0
8:00 222.0 F 6.2
10:00 228.2 F 8.7
12:00 239.2 F 12.4
14:00 239.7 F 10.7
16:00 236.3 F 8.7
18:00 250.1 F 15.4
20:00 261.4 F 21.9
22:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 271.8 F 28.5
2:00 272.4 F 29.0
3:00 271.4 F 28.6
4:00 271.6 F 28.6

Min 271.2 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 227.1 F 0.0
10:00 231.2 F 0.0
15:00 230.4 F 0.0
20:00 228.6 F 0.0
25:00 232.4 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

10/04/2023 04:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____