
Midmark M11 - v1.0.5
Total Cycles: 3574

Name: RITTER2
Sterilizer ID: V2313228
05/01/2024 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.1 F 0.0
2:00 155.2 F 1.4
4:00 156.5 F 4.8
6:00 175.4 F 8.2
8:00 194.3 F 12.6
10:00 216.1 F 6.4
12:00 225.0 F 10.7
14:00 229.1 F 7.7
16:00 238.8 F 12.4
18:00 238.3 F 10.6
20:00 250.9 F 16.8
22:00 261.9 F 23.2
24:00 272.2 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.6
1:00 272.1 F 30.4
2:00 272.8 F 30.3
3:00 272.1 F 29.9
4:00 271.3 F 29.5

Min 271.1 F 29.5
Max 272.9 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 221.0 F 0.0
10:00 209.4 F 0.0
15:00 213.9 F 0.0
20:00 234.6 F 0.0
25:00 229.8 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:36
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:03:05
CYCLE COMPLETE

05/01/2024 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____