

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006715  
11:15 11/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:04  
PRESSURIZING 2:06  
STERILIZING 3:09  
136.4 C 222kPa 3:09  
CYCLE ABORTED 3:35  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 3:35  
-----

Digital Signature #  
06D333310D252226  
-----