
Midmark M11 - v1.0.0
Total Cycles: 3159

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2023 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.9 F 0.0
2:00 185.5 F 0.0
4:00 206.0 F 2.5
6:00 223.6 F 7.7
8:00 236.8 F 12.9
10:00 236.2 F 9.7
12:00 238.7 F 8.6
14:00 242.4 F 10.9
16:00 248.4 F 14.2
18:00 261.6 F 21.6
20:00 272.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.3 F 28.5
2:00 272.1 F 28.6
3:00 272.1 F 28.3
4:00 272.2 F 28.7

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 228.7 F 0.0
10:00 225.0 F 0.0
15:00 225.1 F 0.0
20:00 226.9 F 0.0
25:00 229.1 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 20:16
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

04/25/2023 10:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____