

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1169

Name: RITTER2  
Sterilizer ID: V2313228  
06/07/2022 01:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.1 F 0.0  
2:00 162.0 F 2.5  
4:00 170.3 F 6.6  
6:00 192.9 F 11.8  
8:00 215.6 F 6.2  
10:00 230.3 F 11.3  
12:00 231.3 F 8.3  
14:00 240.7 F 11.4  
16:00 244.5 F 13.2  
18:00 257.0 F 19.9  
20:00 268.1 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.5 F 29.9  
2:00 271.6 F 29.1  
3:00 272.4 F 30.0  
4:00 272.2 F 29.5  
  
Min 271.4 F 29.1  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 216.6 F 0.0  
10:00 226.0 F 0.0  
15:00 228.5 F 0.0  
20:00 228.5 F 0.0  
25:00 226.2 F 0.0  
30:00 222.4 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 21:18  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:53  
CYCLE COMPLETE

06/07/2022 03:18 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02