
Midmark M11 - v1.0.0
Total Cycles: 1755

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2022 01:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.0 F 0.0
2:00 162.9 F 2.2
4:00 194.0 F 5.7
6:00 211.7 F 9.9
8:00 224.5 F 6.2
10:00 230.4 F 9.7
12:00 235.2 F 8.5
14:00 242.4 F 12.1
16:00 239.9 F 10.3
18:00 253.1 F 16.9
20:00 264.6 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.1
2:00 271.4 F 28.3
3:00 272.2 F 29.2
4:00 272.2 F 28.7

Min 271.2 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 228.7 F 0.0
10:00 229.4 F 0.0
15:00 237.7 F 0.0
20:00 231.7 F 0.0
25:00 229.2 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

03/28/2022 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____