

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002082  
11:16 23/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:36  
STERILIZING 3:48  
136.5 C 226kPa 3:48  
136.3 C 226kPa 4:18  
136.4 C 226kPa 4:48  
136.4 C 226kPa 5:18  
136.3 C 227kPa 5:48  
136.4 C 226kPa 6:18  
136.4 C 225kPa 6:48  
136.3 C 224kPa 7:18  
136.3 C 224kPa 7:18  
VENTING 7:18  
AIR DRYING 8:12  
CYCLE COMPLETE 27:06  
-----

Digital Signature #  
C7E83D9211D78F39  
-----