

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1758

Name: RITTER 1  
Sterilizer ID: V2178198  
03/29/2022 07:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.0 F 0.0  
2:00 77.6 F 0.0  
4:00 86.5 F 0.0  
6:00 112.2 F 0.4  
8:00 147.4 F 2.3  
10:00 175.1 F 4.9  
12:00 194.1 F 8.3  
14:00 212.0 F 12.3  
16:00 219.7 F 5.3  
18:00 230.1 F 9.5  
20:00 238.2 F 10.9  
22:00 239.7 F 10.7  
24:00 235.1 F 8.2  
26:00 249.8 F 15.0  
28:00 261.2 F 21.4  
30:00 271.8 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.7 F 28.9  
2:00 271.8 F 28.1  
3:00 271.8 F 28.5  
4:00 272.6 F 28.8

Min 271.3 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 220.0 F 0.0  
10:00 222.8 F 0.0  
15:00 226.1 F 0.0  
20:00 227.4 F 0.0  
25:00 231.9 F 0.0  
30:00 227.2 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 30:38  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:08:27  
CYCLE COMPLETE

03/29/2022 08:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_