
Midmark M11 - v1.0.5
Total Cycles: 1520

Name: RITTER2
Sterilizer ID: V2313228
09/22/2022 06:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 76.6 F 0.0
4:00 91.8 F 0.0
6:00 114.5 F 0.1
8:00 175.9 F 0.3
10:00 202.9 F 3.0
12:00 207.8 F 6.8
14:00 230.5 F 10.7
16:00 231.9 F 7.1
18:00 236.1 F 9.7
20:00 240.0 F 10.2
22:00 243.6 F 12.4
24:00 241.4 F 11.3
26:00 254.2 F 17.8
28:00 265.4 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 273.0 F 29.5
2:00 272.8 F 29.6
3:00 272.7 F 29.6
4:00 272.4 F 29.7

Min 271.0 F 28.6
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 173.7 F 0.0
10:00 176.1 F 0.0
15:00 170.2 F 0.0
20:00 164.9 F 0.0
25:00 163.2 F 0.0
30:00 161.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 29:48
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:04
CYCLE COMPLETE

09/22/2022 07:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____