
Midmark M11 - v1.0.5
Total Cycles: 2095

Name: RITTER2
Sterilizer ID: V2313228
02/27/2023 03:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.0 F 0.0
2:00 181.6 F 4.0
4:00 186.0 F 8.7
6:00 221.4 F 8.2
8:00 229.2 F 9.6
10:00 231.3 F 8.3
12:00 244.1 F 13.5
14:00 243.9 F 12.8
16:00 257.3 F 20.0
18:00 269.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.7 F 29.8
2:00 271.6 F 29.0
3:00 272.1 F 29.9
4:00 272.5 F 29.7

Min 271.1 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 234.5 F 0.0
10:00 228.0 F 0.0
15:00 224.0 F 0.0
20:00 224.8 F 0.0
25:00 231.0 F 0.0
30:00 225.5 F 0.0

Drying Complete

FILLING: 1:34
HEATING: 19:01
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

02/27/2023 04:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____