

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001116  
12:14 09/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:30  
PRESSURIZING 4:30  
STERILIZING 5:49  
136.0 C 220kPa 5:49  
136.2 C 223kPa 6:19  
136.0 C 221kPa 6:49  
136.4 C 225kPa 7:19  
136.4 C 225kPa 7:49  
136.3 C 224kPa 8:19  
136.4 C 224kPa 8:49  
136.1 C 223kPa 9:19  
136.1 C 223kPa 9:19  
VENTING 9:20  
AIR DRYING 10:16  
CYCLE COMPLETE 17:21  
-----

Digital Signature #  
3663DA3CCBBCBDAC  
-----