
Midmark M11 - v1.0.5
Total Cycles: 826

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.6 F 0.0
2:00 182.1 F 3.8
4:00 196.2 F 9.0
6:00 219.3 F 8.0
8:00 230.6 F 10.7
10:00 232.9 F 8.6
12:00 239.0 F 10.2
14:00 248.6 F 14.8
16:00 260.9 F 22.0
18:00 272.3 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 271.5 F 29.2
2:00 272.6 F 29.6
3:00 272.0 F 29.1
4:00 271.1 F 28.7

Min 271.0 F 28.6
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 215.6 F 0.0
10:00 222.7 F 0.0
15:00 223.0 F 0.0
20:00 221.2 F 0.0
25:00 219.8 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:25
CYCLE COMPLETE

02/23/2022 12:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____