

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4316

Name: RITTER 1  
Sterilizer ID: V2178198  
02/21/2024 01:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.1 F 0.0  
2:00 165.5 F 2.5  
4:00 169.1 F 6.0  
6:00 190.5 F 9.9  
8:00 211.3 F 7.5  
10:00 218.2 F 6.7  
12:00 229.4 F 11.0  
14:00 229.8 F 7.0  
16:00 241.5 F 12.5  
18:00 240.4 F 10.8  
20:00 252.5 F 16.8  
22:00 263.2 F 23.0  
24:00 272.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.5 F 29.2  
2:00 271.9 F 28.6  
3:00 271.5 F 28.7  
4:00 272.4 F 29.2

Min 271.3 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 221.4 F 0.0  
10:00 231.1 F 0.0  
15:00 229.6 F 0.0  
20:00 228.8 F 0.0  
25:00 223.2 F 0.0  
30:00 215.9 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 24:21  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:03:27  
CYCLE COMPLETE

02/21/2024 02:31 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_