
Midmark M11 - v1.0.0
Total Cycles: 3360

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2023 02:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.3 F 0.0
2:00 182.8 F 2.4
4:00 201.6 F 7.6
6:00 221.9 F 13.3
8:00 229.8 F 8.8
10:00 236.7 F 10.1
12:00 243.0 F 12.5
14:00 242.1 F 11.5
16:00 255.8 F 18.5
18:00 267.5 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.2
2:00 271.8 F 28.5
3:00 272.0 F 29.0
4:00 272.5 F 29.1

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 214.6 F 0.0
10:00 227.3 F 0.0
15:00 237.4 F 0.0
20:00 229.5 F 0.0
25:00 224.9 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

06/08/2023 03:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____