
Midmark M11 - v1.0.5
Total Cycles: 1891

Name: RITTER2
Sterilizer ID: V2313228
01/04/2023 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.9 F 0.0
2:00 146.7 F 1.9
4:00 156.3 F 4.9
6:00 181.0 F 9.2
8:00 201.3 F 7.7
10:00 225.3 F 8.1
12:00 233.1 F 13.5
14:00 237.9 F 10.7
16:00 236.1 F 9.0
18:00 249.9 F 15.8
20:00 261.9 F 22.7
22:00 272.6 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.1 F 29.5
2:00 272.0 F 30.0
3:00 271.6 F 29.1
4:00 272.2 F 29.8

Min 271.3 F 29.1
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.1 F 0.0
5:00 233.0 F 0.0
10:00 230.4 F 0.0
15:00 221.1 F 0.0
20:00 219.0 F 0.0
25:00 223.0 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

01/04/2023 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____