
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001770
14:27 02/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:40
STERILIZING 3:52
136.5 C 224kPa 3:52
136.4 C 225kPa 4:22
136.3 C 222kPa 4:52
136.3 C 224kPa 5:22
136.3 C 225kPa 5:52
136.4 C 225kPa 6:22
136.4 C 226kPa 6:52
136.4 C 225kPa 7:22
136.4 C 225kPa 7:22
VENTING 7:23
AIR DRYING 8:19
CYCLE COMPLETE 11:35

Digital Signature #
CA7C216AA2F1578F
