
Midmark M11 - v1.0.5
Total Cycles: 3911

Name: RITTER2
Sterilizer ID: V2313228
09/06/2024 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.0 F 0.0
2:00 168.8 F 1.8
4:00 177.6 F 5.2
6:00 199.5 F 9.7
8:00 217.6 F 9.0
10:00 225.1 F 8.4
12:00 237.7 F 13.0
14:00 238.0 F 10.5
16:00 236.0 F 8.8
18:00 249.1 F 15.2
20:00 260.0 F 21.4
22:00 270.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.5 F 29.4
2:00 271.1 F 28.8
3:00 272.3 F 29.5
4:00 271.4 F 28.8

Min 271.0 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 228.1 F 0.0
10:00 200.1 F 0.0
15:00 196.9 F 0.0
20:00 192.6 F 0.0
25:00 188.7 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 22:54
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

09/06/2024 11:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____