

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2950

Name: RITTER 1  
Sterilizer ID: V2178198  
02/27/2023 05:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.2 F 0.0  
2:00 172.0 F 3.5  
4:00 187.2 F 7.5  
6:00 217.4 F 12.9  
8:00 225.5 F 7.4  
10:00 237.4 F 12.5  
12:00 237.2 F 9.9  
14:00 237.7 F 8.4  
16:00 249.7 F 15.1  
18:00 261.7 F 21.9  
20:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.9 F 29.1  
2:00 271.4 F 28.4  
3:00 272.9 F 29.3  
4:00 271.9 F 29.1  
  
Min 271.0 F 28.2  
Max 273.1 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 220.2 F 0.0  
10:00 232.2 F 0.0  
15:00 232.7 F 0.0  
20:00 226.5 F 0.0  
25:00 227.0 F 0.0  
30:00 224.3 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 20:25  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

02/27/2023 06:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_