
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001707
15:51 21/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:36
STERILIZING 3:46
136.5 C 224kPa 3:46
136.3 C 224kPa 4:16
136.2 C 221kPa 4:46
136.2 C 222kPa 5:16
136.4 C 223kPa 5:46
136.4 C 225kPa 6:16
136.2 C 222kPa 6:46
136.2 C 222kPa 7:16
136.2 C 222kPa 7:16
VENTING 7:16
AIR DRYING 8:04
CYCLE COMPLETE 8:13

Digital Signature #
5106CFFE5BB44B73
