
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001556
8:52 11/04/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:11
PRESSURIZING 7:08
STERILIZING 8:47
136.0 C 220kPa 8:47
136.5 C 221kPa 9:17
136.2 C 220kPa 9:47
136.1 C 219kPa 10:17
135.8 C 219kPa 10:47
136.4 C 220kPa 11:17
136.4 C 222kPa 11:47
136.1 C 220kPa 12:17
VENTING 12:17
AIR DRYING 13:24
CYCLE COMPLETE 32:18

Digital Signature #
AA56270E2FA0DD29
