

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5189

Name: RITTER 1  
Sterilizer ID: V2178198  
12/13/2024 04:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.6 F 0.0  
2:00 167.4 F 0.2  
4:00 194.8 F 2.7  
6:00 215.3 F 7.5  
8:00 230.5 F 12.8  
10:00 232.3 F 8.4  
12:00 242.8 F 11.8  
14:00 243.3 F 11.8  
16:00 240.9 F 10.3  
18:00 254.6 F 17.2  
20:00 266.2 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.7  
1:00 272.7 F 28.3  
2:00 271.6 F 27.7  
3:00 272.8 F 28.4  
4:00 271.8 F 27.6

Min 271.6 F 27.6  
Max 272.9 F 28.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 217.4 F 0.0  
10:00 212.9 F 0.0  
15:00 204.9 F 0.0  
20:00 221.9 F 0.0  
25:00 245.0 F 0.0  
30:00 239.2 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 21:37  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:59:58  
CYCLE COMPLETE

12/13/2024 05:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_