
Midmark M11 - v1.0.0
Total Cycles: 806

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2021 09:22 AM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.4 F 0.0
2:00 162.3 F 2.2
4:00 182.1 F 5.2
6:00 211.9 F 1.6
8:00 223.7 F 5.1
10:00 224.0 F 4.0
12:00 231.9 F 6.8
14:00 233.1 F 7.1
16:00 245.6 F 12.7

STERILIZING
mm:ss Degrees PSI
0:00 251.5 F 15.7
1:00 252.5 F 16.1
2:00 251.8 F 16.0
3:00 252.4 F 16.0
4:00 251.4 F 15.7
5:00 252.5 F 16.3
6:00 252.3 F 16.0
7:00 251.4 F 15.6
8:00 252.2 F 16.2
9:00 252.5 F 16.2
10:00 252.3 F 16.0
11:00 251.7 F 15.8
12:00 251.4 F 15.7
13:00 252.2 F 16.2
14:00 252.6 F 16.3
15:00 252.3 F 16.2
16:00 252.3 F 16.0
17:00 252.0 F 15.9
18:00 251.5 F 15.7
19:00 251.3 F 15.7
20:00 251.8 F 16.1
21:00 252.6 F 16.4
22:00 252.6 F 16.3
23:00 252.5 F 16.2
24:00 252.3 F 16.1
25:00 252.1 F 16.0
26:00 251.8 F 15.8
27:00 251.5 F 15.7
28:00 251.4 F 15.7
29:00 251.7 F 16.0
30:00 252.4 F 16.3

Min 251.3 F 15.6
Max 252.8 F 16.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 227.8 F 0.0
10:00 222.6 F 0.0
15:00 226.6 F 0.0
20:00 213.7 F 0.0
25:00 166.1 F 0.0
Dry Cycle Aborted

FILLING: 1:33
HEATING: 17:39
STERILIZING: 30:00

VENTING: 3:01
DRYING: 28:52
TOTAL CYCLE: 01:21:05

04/26/2021 10:43 AM
ACCEPTED: X
REJECTED:
Operator: U01