
Midmark M11 - v1.0.5
Total Cycles: 4158

Name: RITTER2
Sterilizer ID: V2313228
11/26/2024 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.2 F 0.0
2:00 103.3 F 0.0
4:00 116.9 F 0.0
6:00 165.9 F 0.2
8:00 188.7 F 0.6
10:00 203.6 F 3.0
12:00 215.3 F 6.4
14:00 227.5 F 10.4
16:00 231.6 F 8.7
18:00 240.2 F 11.9
20:00 238.7 F 10.5
22:00 237.7 F 9.7
24:00 251.5 F 16.4
26:00 263.0 F 23.1
28:00 272.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.3
2:00 271.7 F 29.0
3:00 272.5 F 29.2
4:00 271.7 F 29.0

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 217.1 F 0.0
10:00 219.8 F 0.0
15:00 221.9 F 0.0
20:00 218.9 F 0.0
25:00 220.8 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:11
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:06:52
CYCLE COMPLETE

11/26/2024 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____