

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4184

Name: RITTER2  
Sterilizer ID: V2313228  
12/04/2024 03:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.8 F 0.0  
2:00 144.2 F 0.4  
4:00 162.5 F 2.8  
6:00 189.1 F 6.9  
8:00 213.4 F 11.7  
10:00 221.2 F 7.1  
12:00 232.9 F 12.5  
14:00 235.8 F 10.2  
16:00 235.8 F 8.7  
18:00 249.5 F 15.7  
20:00 261.2 F 22.5  
22:00 272.5 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.9  
1:00 271.9 F 29.8  
2:00 272.3 F 29.8  
3:00 272.9 F 30.0  
4:00 272.7 F 30.3

Min 271.6 F 29.7  
Max 273.3 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 232.5 F 0.0  
10:00 238.6 F 0.0  
15:00 216.6 F 0.0  
20:00 187.2 F 0.0  
25:00 180.6 F 0.0  
30:00 178.5 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 23:08  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:02:00  
CYCLE COMPLETE

12/04/2024 04:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_