
Midmark M11 - v1.0.0
Total Cycles: 4509

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2024 10:35 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 173.0 F 3.4
4:00 180.3 F 7.6
6:00 213.7 F 12.0
8:00 222.6 F 5.7
10:00 233.7 F 10.7
12:00 230.4 F 7.0
14:00 244.2 F 13.0
16:00 242.1 F 11.4
18:00 254.9 F 17.8
20:00 266.1 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.8
2:00 271.2 F 28.2
3:00 272.4 F 28.9

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 224.3 F 0.0
10:00 214.4 F 0.0
15:00 235.8 F 0.0
20:00 218.3 F 0.0
25:00 217.1 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:39
STERILIZING: 3:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

05/01/2024 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____