
Midmark M11 - v1.0.5
Total Cycles: 170

Name: RITTER2
Sterilizer ID: V2313228
07/07/2021 06:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 75.4 F 0.1
4:00 77.9 F 0.9
6:00 112.3 F 3.0
8:00 140.0 F 5.8
10:00 166.6 F 9.4
12:00 188.4 F 13.3
14:00 201.7 F 5.1
16:00 214.7 F 8.6
18:00 226.4 F 12.5
20:00 232.7 F 8.8
22:00 240.5 F 11.9
24:00 242.7 F 12.1
26:00 254.1 F 18.1
28:00 264.5 F 24.4
30:00 270.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 272.4 F 29.7
2:00 272.5 F 30.1
3:00 272.2 F 30.1
4:00 271.9 F 29.5

Min 271.5 F 29.5
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.6 F 0.0
5:00 219.0 F 0.0
10:00 236.5 F 0.0
15:00 201.6 F 0.0
20:00 203.6 F 0.0
25:00 208.1 F 0.0
30:00 202.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:50
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

07/07/2021 07:57 AM
ACCEPTED: X
REJECTED:
Operator: U01