
Midmark M11 - v1.0.5
Total Cycles: 1882

Name: RITTER2
Sterilizer ID: V2313228
12/22/2022 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.3 F 0.0
2:00 196.5 F 3.4
4:00 208.6 F 8.8
6:00 225.8 F 8.1
8:00 233.8 F 10.6
10:00 237.4 F 8.3
12:00 241.4 F 10.4
14:00 247.9 F 14.6
16:00 260.8 F 21.8
18:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.1 F 29.1
2:00 271.5 F 29.3
3:00 272.0 F 29.0
4:00 271.4 F 29.2

Min 271.1 F 28.6
Max 272.3 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 207.0 F 0.0
10:00 223.3 F 0.0
15:00 230.4 F 0.0
20:00 230.7 F 0.0
25:00 220.1 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

12/22/2022 02:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____