
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000772
14:28 28/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:08
STERILIZING 4:17
136.0 C 230kPa 4:17
136.4 C 232kPa 4:47
136.4 C 230kPa 5:17
136.3 C 232kPa 5:47
136.3 C 232kPa 6:17
136.4 C 233kPa 6:47
136.4 C 233kPa 7:17
136.3 C 232kPa 7:47
136.3 C 232kPa 7:47
VENTING 7:47
AIR DRYING 8:42
CYCLE COMPLETE 8:45

Digital Signature #
6416A53C89748897
