
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001042
12:35 20/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:54
STERILIZING 5:17
136.4 C 222kPa 5:17
136.3 C 223kPa 5:47
136.3 C 223kPa 6:17
136.3 C 222kPa 6:47
136.4 C 225kPa 7:17
136.4 C 224kPa 7:47
136.2 C 222kPa 8:17
136.3 C 223kPa 8:47
136.3 C 223kPa 8:47
VENTING 8:47
AIR DRYING 9:44
CYCLE COMPLETE 24:45

Digital Signature #
2ED7C91065C185E7
