

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001622  
15:16 11/25/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:27  
PRESSURIZING 7:26  
STERILIZING 9:02  
136.4 C 221kPa 9:02  
136.2 C 221kPa 9:32  
136.1 C 221kPa 10:02  
136.4 C 222kPa 10:32  
136.5 C 223kPa 11:02  
136.4 C 223kPa 11:32  
136.3 C 222kPa 12:02  
136.2 C 219kPa 12:32  
VENTING 12:32  
AIR DRYING 13:42  
CYCLE COMPLETE 43:42  
-----

Digital Signature #  
C72B62DDFC84A907  
-----