
Midmark M11 - v1.0.0
Total Cycles: 1961

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2022 06:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 75.4 F 0.0
4:00 85.1 F 0.1
6:00 125.7 F 2.3
8:00 153.4 F 5.0
10:00 172.5 F 8.2
12:00 191.7 F 11.8
14:00 208.6 F 2.8
16:00 217.3 F 6.4
18:00 227.0 F 9.9
20:00 235.5 F 10.9
22:00 236.7 F 10.0
24:00 236.9 F 8.8
26:00 245.9 F 13.4
28:00 257.1 F 19.3
30:00 267.2 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.5 F 29.3
2:00 272.1 F 28.7
3:00 271.5 F 28.7
4:00 272.5 F 29.3

Min 271.2 F 28.3
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 204.4 F 0.0
10:00 200.5 F 0.0
15:00 193.1 F 0.0
20:00 217.3 F 0.0
25:00 225.0 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:33
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:09:31
CYCLE COMPLETE

05/27/2022 08:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____