

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5241

Name: RITTER 1  
Sterilizer ID: V2178198  
01/14/2025 11:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.2 F 0.0  
2:00 69.8 F 0.0  
4:00 84.2 F 0.0  
6:00 113.2 F 0.0  
8:00 142.5 F 0.1  
10:00 170.4 F 0.2  
12:00 188.1 F 0.5  
14:00 198.3 F 0.9  
16:00 212.1 F 4.0  
18:00 222.7 F 6.6  
20:00 229.8 F 9.7  
22:00 239.1 F 13.4  
24:00 237.0 F 9.8  
26:00 240.1 F 10.1  
28:00 245.1 F 12.8  
30:00 242.3 F 11.3  
32:00 253.4 F 16.8  
34:00 263.1 F 22.6  
36:00 272.1 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.1 F 28.5  
2:00 272.7 F 28.9  
3:00 272.6 F 28.7  
4:00 272.2 F 28.4

Min 271.6 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.0 F 0.0  
5:00 198.5 F 0.0  
10:00 208.8 F 0.0  
15:00 230.4 F 0.0  
20:00 234.4 F 0.0  
25:00 230.0 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 36:38  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:15:12  
CYCLE COMPLETE

01/14/2025 12:35 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_