
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002371
8:49 13/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:43
STERILIZING 3:46
136.4 C 223kPa 3:46
136.4 C 225kPa 4:16
136.4 C 225kPa 4:46
136.4 C 224kPa 5:16
136.1 C 222kPa 5:46
136.4 C 223kPa 6:16
136.4 C 225kPa 6:46
136.4 C 224kPa 7:16
136.4 C 224kPa 7:16
VENTING 7:16
AIR DRYING 8:13
CYCLE COMPLETE 38:13

Digital Signature #
4381FE0DA32193B0
