

-----  
Midmark M11 - v1.0.5  
Total Cycles: 679

Name: RITTER2  
Sterilizer ID: V2313228  
01/11/2022 09:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.1 F 0.0  
2:00 77.5 F 0.0  
4:00 94.0 F 0.0  
6:00 108.7 F 0.7  
8:00 136.7 F 3.4  
10:00 162.7 F 6.7  
12:00 182.6 F 10.3  
14:00 200.5 F 7.3  
16:00 214.4 F 6.6  
18:00 224.9 F 10.5  
20:00 230.8 F 8.8  
22:00 237.8 F 11.6  
24:00 236.3 F 9.4  
26:00 248.7 F 15.3  
28:00 259.3 F 21.1  
30:00 269.1 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.7 F 29.9  
2:00 271.7 F 29.1  
3:00 272.0 F 30.0  
4:00 272.5 F 29.6

Min 271.4 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.6 F 0.0  
5:00 198.7 F 0.0  
10:00 212.4 F 0.0  
15:00 209.3 F 0.0  
20:00 206.1 F 0.0  
25:00 205.5 F 0.0  
30:00 211.3 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 31:15  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:09:53  
CYCLE COMPLETE

01/11/2022 10:25 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_