

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5237

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2025 05:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.1 F 0.0  
2:00 148.6 F 0.2  
4:00 175.7 F 2.5  
6:00 200.0 F 6.9  
8:00 211.4 F 11.5  
10:00 225.2 F 5.7  
12:00 232.6 F 10.5  
14:00 233.5 F 7.7  
16:00 245.1 F 13.0  
18:00 242.6 F 11.5  
20:00 255.5 F 18.0  
22:00 266.7 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 271.6 F 28.1  
2:00 271.8 F 28.3  
3:00 272.1 F 28.5  
4:00 272.4 F 28.8

Min 271.4 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 199.2 F 0.0  
10:00 233.9 F 0.0  
15:00 221.4 F 0.0  
20:00 216.3 F 0.0  
25:00 231.4 F 0.0  
30:00 210.4 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 23:26  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:01:54  
CYCLE COMPLETE

01/10/2025 06:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_