
Midmark M11 - v1.0.0
Total Cycles: 5160

Name: RITTER 1
Sterilizer ID: V2178198
12/04/2024 06:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.3 F 0.0
2:00 171.6 F 0.3
4:00 206.5 F 1.1
6:00 219.7 F 4.7
8:00 228.5 F 9.0
10:00 238.5 F 11.8
12:00 240.0 F 10.8
14:00 236.0 F 8.8
16:00 239.2 F 9.6
18:00 249.9 F 15.0
20:00 262.2 F 22.0
22:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.8
2:00 271.7 F 28.1
3:00 272.5 F 28.9
4:00 272.3 F 28.6

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 233.7 F 0.0
10:00 221.6 F 0.0
15:00 199.7 F 0.0
20:00 186.9 F 0.0
25:00 184.4 F 0.0
30:00 182.3 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:25
CYCLE COMPLETE

12/04/2024 07:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____