
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000360
13:41 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:12
STERILIZING 3:12
136.0 C 226kPa 3:12
136.4 C 229kPa 3:42
136.4 C 233kPa 4:12
136.4 C 233kPa 4:42
136.4 C 233kPa 5:12
136.4 C 233kPa 5:42
136.2 C 232kPa 6:12
136.4 C 233kPa 6:42
136.4 C 233kPa 6:42
VENTING 6:43
AIR DRYING 7:33
CYCLE COMPLETE 7:47

Digital Signature #
79F607F30E9880CC
