
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007027
13:14 01/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:28
STERILIZING 3:27
136.5 C 223kPa 3:27
136.4 C 225kPa 3:57
136.1 C 221kPa 4:27
136.4 C 224kPa 4:57
136.4 C 224kPa 5:27
136.4 C 223kPa 5:57
136.4 C 222kPa 6:27
136.0 C 218kPa 6:57
136.0 C 218kPa 6:57
VENTING 6:57
AIR DRYING 7:49
CYCLE COMPLETE 25:55

Digital Signature #
97C56A8C2A2DF470
