
Midmark M11 - v1.0.5
Total Cycles: 3315

Name: RITTER2
Sterilizer ID: V2313228
01/24/2024 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 153.3 F 1.7
4:00 162.4 F 4.7
6:00 185.5 F 8.8
8:00 201.6 F 13.1
10:00 222.0 F 7.0
12:00 235.2 F 11.4
14:00 234.0 F 8.2
16:00 242.2 F 11.7
18:00 245.4 F 13.0
20:00 257.4 F 19.6
22:00 268.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 271.8 F 29.0
2:00 272.4 F 29.3
3:00 272.3 F 29.5
4:00 271.6 F 29.6

Min 271.0 F 29.0
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.0 F 0.0
5:00 192.9 F 0.0
10:00 192.6 F 0.0
15:00 195.4 F 0.0
20:00 195.6 F 0.0
25:00 195.6 F 0.0
30:00 195.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:48
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:02:25
CYCLE COMPLETE

01/24/2024 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____