
Midmark M11 - v1.0.0
Total Cycles: 3451

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2023 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 187.7 F 2.3
4:00 198.8 F 6.3
6:00 215.3 F 11.4
8:00 222.6 F 6.0
10:00 234.7 F 10.9
12:00 233.2 F 8.0
14:00 243.7 F 12.0
16:00 245.2 F 12.8
18:00 257.2 F 19.2
20:00 267.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.0
2:00 271.9 F 28.4
3:00 271.9 F 28.7
4:00 272.7 F 29.0

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 208.8 F 0.0
10:00 212.4 F 0.0
15:00 215.5 F 0.0
20:00 215.3 F 0.0
25:00 212.7 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:00:30
CYCLE COMPLETE

06/27/2023 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____