
Midmark M11 - v1.0.0
Total Cycles: 623

Name: MIDMARK
Sterilizer ID: V2178198
01/25/2021 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.8 F 0.0
2:00 93.1 F 0.0
4:00 149.6 F 0.0
6:00 145.0 F 0.2
8:00 180.7 F 2.4
10:00 199.3 F 6.0
12:00 221.0 F 9.8
14:00 231.2 F 9.7
16:00 232.7 F 9.2
18:00 238.4 F 10.2
20:00 243.3 F 11.9
22:00 240.2 F 10.1
24:00 253.7 F 16.8
26:00 265.2 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.6 F 28.5
2:00 272.4 F 28.4
3:00 272.2 F 28.3
4:00 272.2 F 28.4

Min 270.9 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 234.6 F 0.0
10:00 223.9 F 0.0
15:00 224.1 F 0.0
20:00 231.0 F 0.0
25:00 229.6 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:37
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:05:24
CYCLE COMPLETE

01/25/2021 11:38 AM
ACCEPTED: X
REJECTED:
Operator: U01