
Midmark M11 - v1.0.5
Total Cycles: 2045

Name: RITTER2
Sterilizer ID: V2313228
02/14/2023 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.8 F 0.0
2:00 181.9 F 1.7
4:00 188.9 F 5.8
6:00 210.0 F 11.0
8:00 221.7 F 6.7
10:00 235.8 F 12.5
12:00 236.9 F 10.4
14:00 236.4 F 9.2
16:00 251.2 F 16.4
18:00 263.1 F 23.5
20:00 271.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.7
2:00 271.7 F 29.0
3:00 272.1 F 29.8
4:00 272.4 F 29.6

Min 271.3 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 226.3 F 0.0
10:00 236.6 F 0.0
15:00 228.2 F 0.0
20:00 225.4 F 0.0
25:00 225.8 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

02/14/2023 10:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____