
Midmark M11 - v1.0.0
Total Cycles: 5354

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2025 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 166.1 F 3.0
4:00 178.0 F 7.0
6:00 197.3 F 11.5
8:00 217.0 F 4.2
10:00 227.6 F 8.9
12:00 237.6 F 11.1
14:00 239.7 F 10.7
16:00 236.2 F 8.6
18:00 250.8 F 15.4
20:00 262.8 F 22.2
22:00 272.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.6 F 28.9
2:00 272.4 F 28.4
3:00 271.3 F 27.9
4:00 272.3 F 28.8

Min 271.3 F 27.8
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 183.3 F 0.0
10:00 193.8 F 0.0
15:00 205.5 F 0.0
20:00 207.5 F 0.0
25:00 209.5 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

02/21/2025 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____