
Midmark M11 - v1.0.0
Total Cycles: 4036

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2023 02:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.4 F 0.0
2:00 109.9 F 1.1
4:00 119.9 F 3.4
6:00 147.0 F 6.4
8:00 197.3 F 10.3
10:00 208.9 F 4.4
12:00 212.8 F 6.6
14:00 224.5 F 10.7
16:00 228.5 F 7.3
18:00 239.3 F 11.6
20:00 236.8 F 9.5
22:00 249.8 F 15.5
24:00 261.1 F 21.8
26:00 271.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.3
2:00 271.6 F 28.6
3:00 271.8 F 29.1
4:00 272.3 F 29.1

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.7 F 0.0
5:00 215.8 F 0.0
10:00 224.5 F 0.0
15:00 228.9 F 0.0
20:00 229.6 F 0.0
25:00 229.6 F 0.0
30:00 228.9 F 0.0

Drying Complete

FILLING: 1:14
HEATING: 26:40
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:05:30
CYCLE COMPLETE

11/17/2023 04:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____