
Midmark M11 - v1.0.0
Total Cycles: 5076

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2024 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.0 F 0.0
2:00 70.2 F 0.0
4:00 90.7 F 0.0
6:00 122.4 F 0.0
8:00 151.2 F 0.2
10:00 177.1 F 0.7
12:00 198.2 F 2.2
14:00 212.9 F 5.2
16:00 220.6 F 8.6
18:00 230.2 F 12.3
20:00 230.4 F 7.6
22:00 241.7 F 12.5
24:00 239.0 F 10.2
26:00 235.9 F 8.3
28:00 249.9 F 15.1
30:00 261.3 F 21.6
32:00 271.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 271.6 F 28.2
2:00 271.6 F 28.2
3:00 271.7 F 28.1
4:00 271.8 F 28.2

Min 271.6 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 226.1 F 0.0
10:00 227.5 F 0.0
15:00 231.2 F 0.0
20:00 230.5 F 0.0
25:00 223.4 F 0.0
30:00 218.9 F 0.0

Drying Complete

FILLING: 0:44
HEATING: 32:26
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:32
CYCLE COMPLETE

11/11/2024 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____