
Midmark M11 - v1.0.0
Total Cycles: 5347

Name: RITTER 1
Sterilizer ID: V2178198
02/20/2025 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 71.3 F 0.3
4:00 89.0 F 1.9
6:00 132.4 F 4.3
8:00 153.6 F 7.0
10:00 173.9 F 10.3
12:00 185.7 F 6.5
14:00 206.1 F 4.2
16:00 219.1 F 7.7
18:00 229.5 F 11.4
20:00 226.5 F 6.2
22:00 239.7 F 11.3
24:00 236.2 F 8.7
26:00 249.3 F 14.9
28:00 260.2 F 21.0
30:00 270.4 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.8 F 29.0
2:00 271.7 F 28.2
3:00 272.6 F 29.1
4:00 272.1 F 28.5

Min 271.4 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 187.2 F 0.0
10:00 183.5 F 0.0
15:00 211.5 F 0.0
20:00 232.4 F 0.0
25:00 226.2 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:53
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:09:15
CYCLE COMPLETE

02/20/2025 11:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____