
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000277
13:39 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:39
STERILIZING 5:09
136.4 C 233kPa 5:09
136.3 C 233kPa 5:39
136.4 C 233kPa 6:09
136.3 C 232kPa 6:39
136.4 C 233kPa 7:09
136.4 C 234kPa 7:39
136.2 C 233kPa 8:09
136.3 C 233kPa 8:39
136.3 C 233kPa 8:39
VENTING 8:39
AIR DRYING 9:44
CYCLE COMPLETE 24:44

Digital Signature #
7CEE8B47D56FA689
